



## *EC Declaration of Conformity*

Manufacturer:

Name: Acro Biotech, Inc.

Address: 9500, 7th str., Unit M, Rancho Cucamonga, CA 91730, USA

European Representative:

Name: MedNet GmbH

Address: Borkstrasse 10, 48163 Muenster , Germany

Product Name: COVID-19 (Novel Coronavirus) Antigen Rapid Test (Oral Fluid)

Model: Cassette

Classification: Other Device of IVDD 98/79/EC

Conformity Assessment Route: IVDD 98/79/EC Annex III (excluding point 6)

EDMA Code: 15 70 90 08 00

We, Acro Biotech, Inc., herewith declare that we are exclusively responsible for this declaration of conformity. We herewith declare that the above mentioned products meet the transposition into national law, the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer.

### DIRECTIVES

General applicable directives:

DIRECTIVE 98/79/EC OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 27 October 1998 on in vitro diagnostic medical devices

Standard Applied: EN ISO 13485:2016, EN ISO 14971:2012, EN 13975:2003, EN ISO 18113-1:2011, EN ISO 18113-2:2011, EN 13612:2002/AC:2002, EN ISO 17511:2003, EN ISO 23640:2015, EN 13641:2002, EN ISO 15223-1:2016

Place, Date of Issue: in Rancho Cucamonga on 12/11/2020

Signature: \_\_\_\_\_

Name: Joseph Fan

Position: President

**ACRO BIOTECH, INC**

**ACRO BIOTECH, Inc.**

9500 Seventh Street,  
Unit M, Rancho Cucamonga, CA 91730, U.S.A.  
Tel: +1 (909) 466-6892 info@acrobiotech.com  
www.acrobiotech.com